Atmos. Meas. Tech., 9, 1313–1324, 2016 www.atmos-meas-tech.net/9/1313/2016/ doi:10.5194/amt-9-1313-2016 © Author(s) 2016. CC Attribution 3.0 License.





Re-evaluating the Frankfurt isothermal static diffusion chamber for ice nucleation

Jann Schrod¹, Anja Danielczok¹, Daniel Weber¹, Martin Ebert², Erik S. Thomson³, and Heinz G. Bingemer¹

¹Institute for Atmospheric and Environmental Sciences, J. W. Goethe University, Frankfurt am Main, 60438 Frankfurt, Germany
²Institute for Applied Geosciences, Technical University of Darmstadt, 64287 Darmstadt, Germany
³Department of Chemistry and Molecular Biology, Atmospheric Science, University of Gothenburg, 41296 Gothenburg, Sweden

Correspondence to: Erik S. Thomson (erik.thomson@chem.gu.se)

Received: 11 November 2015 – Published in Atmos. Meas. Tech. Discuss.: 1 December 2015 Revised: 11 March 2016 – Accepted: 15 March 2016 – Published: 30 March 2016

Abstract. Recently significant advances have been made in the collection, detection and characterization of ice nucleating particles (INPs). Ice nuclei are particles that facilitate the heterogeneous formation of ice within the atmospheric aerosol by lowering the free energy barrier to spontaneous nucleation and growth of ice from atmospheric water and/or vapor. The Frankfurt isostatic diffusion chamber (FRankfurt Ice nucleation Deposition freezinG Experiment: FRIDGE) is an INP collection and offline detection system that has become widely deployed and shows additional potential for ambient measurements. Since its initial development FRIDGE has gone through several iterations and improvements. Here we describe improvements that have been made in the collection and analysis techniques. We detail the uncertainties inherent in the measurement method and suggest a systematic method of error analysis for FRIDGE measurements. Thus what is presented herein should serve as a foundation for the dissemination of all current and future measurements using FRIDGE instrumentation.

1 Introduction

Presently significant scientific resources are focused on identifying and categorizing what types of particles within the atmosphere are active ice nuclei. Ice nucleating particles (INPs) are particles suspended within the atmosphere that lower the free-energy barrier that exists to spontaneous nucleation. Thus INPs can be seen as the low-temperature variant of cloud condensation nuclei (CCN), in that they assist in the nucleation of ice in the atmosphere. Almost any type of solid particle can help to nucleate ice in the atmosphere, but it remains an open question what types of particles are best able to lower the barrier to heterogeneous nucleation. Many aerosol particles such as mineral dust and primary biological particles have been identified as ice nucleators, yet the measured abundance of such INPs does not agree with observed macroscopic features of clouds, wherein a host of complicated and dynamic interactions lead to ice particles (DeMott et al., 2011). Although some of the mechanisms of ice multiplication are known (Hallett and Mossop, 1974), it remains a significant scientific hurdle to identify the substances and processes that control the formation of cloud ice and precipitation.

The FRankfurt Ice nucleation Deposition freezinG Experiment (FRIDGE) is a system that pairs electrostatic precipitation of particles onto Si wafers in a collection unit with an isostatic diffusion chamber for the activation, growth and optical detection of ice on ice nucleating particles. The instrument and technique have been previously discussed (Bundke et al., 2008; Klein et al., 2010a); since the original development, new multiple-sampling (Sect. 2 and Schrod et al., 2013) and measurement chamber units (Jiang et al., 2015; Ardon-Dryer and Levin, 2014) have been built. As with any of the current measurement devices used to characterize ice nucleation, FRIDGE has strengths and weaknesses. Its principle strengths are the ease of aerosol collection and storage, which also allow for single-particle electron microscopy analysis. In addition to the automated and remotely programmable collector described in Sect. 2, lightweight collection units have been manufactured that can be deployed onto unmanned aerial vehicles (Born et al., 2012; Lange et al., 2013). Sample wafers can be transported and stored easily for laboratory-based analyses without any strict procedural precautions. A systematic study of wafers exposed to varying storage times and conditions shows little variation in resulting INP counts as described in Sect. 5.1. Disadvantages of FRIDGE that require consideration include the possible evaporation of volatile aerosol constituents during the analysis of samples under medium vacuum conditions, 10^{-1} to 10^2 Pa. Also, the analysis is time intensive and limited to deposition and condensation/immersion mode freezing behavior. Furthermore, after analysis the wafers collected in the sampling process must be prepared for reuse in a laborintensive multi-stage cleaning process. Thus the ratio of sampling time to processing time is one area of focus for continuous incremental improvement.

Since the FRIDGE instrument descriptions were first published by Bundke et al. (2008) and Klein et al. (2010a), there have been significant improvements in wafer analysis and the general understanding of results. Here we summarize recent experimental improvements made to FRIDGE, including a description of the automated PEAC7 (programmable electrostatic aerosol collector) sample collector, and report on errors that are contained in some of the previously published data from the FRIDGE method. The work includes a detailed analysis of these errors and proposes solutions for the FRIDGE user community. Furthermore, a detailed description of measurement uncertainties that must be considered when reporting FRIDGE measurement data is included. The methodological refinements come primarily in three broad areas: (i) improved experimental accuracy for determining nucleated ice, (ii) a clear and systematic understanding of particle and mass losses during the collection process and (iii) a clear quantitative procedure for assessing the measurement uncertainty inherent in reported results.

The existing FRIDGE analysis chambers are all based on the Frankfurt design but have been used in various operating modes (Klein et al., 2010a; Jiang et al., 2015; Ardon-Dryer and Levin, 2014). In particular the Tel Aviv group have employed the FRIDGE chamber to study immersion mode freezing using droplet assays (Ardon-Dryer et al., 2011; Ardon-Dryer and Levin, 2014). Herein we focus on utilizing the FRIDGE technique with ice formation proceeding from a vapor-saturated environment. Thus it should be clear that, although the measurements we discuss at vapor pressures exceeding water saturation can encompass immersion mode freezing, they do not involve freezing in macroscopic droplets. A useful discussion of the nuances pertaining to the pathways of freezing nucleation can be found in Vali et al. (2015) and the associated public discussion materials. The simple design and operation of the FRIDGE collectors make the method uniquely suited to use as a monitoring tool. In this regard FRIDGE has been identified as an important resource in an effort to establish a global network of INP measurement data (Ansmann et al., 2014). Thus it is important that current and future results be understood in the context of the method's systematic error.

2 Automated and remotely controlled multiple-sampling unit

A programmable electrostatic aerosol collector (PEAC7) has been developed for the automated sequential sampling of up to seven wafers. The PEAC7 is based on the design of the single wafer electrostatic aerosol collector as described in Klein et al. (2010a, b) but adds a rotatable wafer tray that includes seven sampling slots. In Fig. 1 a schematic and photograph of the unit are shown side by side. The main body of the sampler made from anodized aluminum consists of a cylindrical housing that encloses a concentric, rotatable plate. The plate rotation is powered by a stepper motor that can position each of seven machined wafer cavities in the sampling position below the charging unit. The charging unit and air intake are integrated into the top plate, which is secured to the underlying main body with a series of set screws during operation and can be removed to exchange the substrate wafers. When secured, an O-ring hermetically seals the top and bottom save for the inlet and outlet of the flow system. The aerosol being sampled flows through the unit at a set rate, while within the sampling chamber particles are negatively charged by collision with corona discharge electrons and are electrostatically precipitated onto the grounded Si wafer underneath (the charging unit). The glass housing of the aerosol inlet and charging unit are clearly visible in Fig. 1. The stepper motor uses a neodymium magnet to calibrate the position of the disc, while a pump, a rotameter, a flow controller and a highvoltage generator enable the sample collection and electrostatic precipitation. The entire unit is mounted in a standard 19 in. rack-mount case and uses a standard IEC 60320 C14 power inlet.

The PEAC7 can be programmed and directly controlled from a digital front panel or vis-á-vis serial communication with a computer through a RS-232 serial port. The unit is designed to allow for automated sampling using programmable date, time, wafer number (1–7), flow rate and sampling duration. The computer link allows for remote control and thus monitoring and reprogramming of scheduled sampling tasks. A number of benefits arise from the use of automated PEAC7 sample collection, including the obvious advantages of reduced manual intervention. Most importantly the process of collecting regular measurements for long-term monitoring is highly simplified. This makes the PEAC7 uniquely suited to unstaffed and/or isolated stations that might be included in regional or global networks for INP monitoring.



Figure 1. Picture of the PEAC7 multiple-sampling unit with a schematic of the charging unit inset. For the photograph the top plate containing the charging unit has been removed and offset to reveal the underlying rotatable disc with seven sample wafers mounted into machined slots.

3 Image analysis

The FRIDGE instrument relies on automated image analysis to identify and count nucleated ice particles on the collection substrates. Experiments demonstrate that the process of ice identification and INP counting are sensitive to illumination, camera resolution and image contrast, and that the INP identification procedure used until 2012 introduced asystematic errors. Thus, FRIDGE data generated between 2008 and 2012 (data appear in the following publications: Bundke et al., 2008; Klein et al., 2010a, b; Bingemer et al., 2012; Niemand et al., 2012) have been identified to include erroneous measurements, resulting from the miscounting of liquid droplets as ice. Unfortunately, the full extent of the error is variable and depends upon the sampled aerosol, meaning that the entirety of data collected from the stated time period and reported in the cited publications must be re-evaluated. Likewise, any conclusions and proposed hypotheses based on INPs counted by FRIDGE from those cited publications must be re-examined.

The experimental procedure and image processing using FRIDGE was previously summarized in Klein et al. (2010a). In this work we review details of the system where previous mistakes were made and clarify how those mistakes have been (or should be) corrected.

A measurement is begun by allowing water vapor to diffuse from the water vapor source, which is a 1.25 L temperature-controlled glass vessel coated with ice, into the evacuated FRIDGE sample cell that encloses the wafer substrate on top of a cold stage, and ice grows on activated INPs. The sample cells are analogous to environmentally controlled microscope stages with the substrate wafers replacing microscope slides. The chamber pressure is monitored to preclude vapor depletion from limiting nucleation, and a CCD camera $(2/3'' \text{CCD} \ge 5 \text{ megapixels}, 1 \text{ pixel} \approx 400 \,\mu\text{m}^2)$ is used to monitor and record the sample substrates. LabView software is used to download images and detect changes in brightness on the aerosol collection wafer surface by comparing realtime images with a reference image taken prior to the introduction of vapor. If more than $ST \ge 30$ (previously $ST \sim 4$) adjacent pixels, where ST is used to indicate a size threshold, are bright enough to exceed a brightness threshold \mathcal{BT} , the area is counted as an ice crystal. Like ST the currently utilized value for $\mathcal{BT} = 30$ exceeds the previously used parameters $\mathcal{BT} \sim 6$ by nearly an order of magnitude. Each pixel domain which exceeds these critical values is counted as an ice crystal and is assumed to originate from a single INP. The FRIDGE chamber has generally been operated in a water vapor regime that is supersaturated with respect to ice but subsaturated with respect to liquid water. This is done to prevent the miscounting of water droplets as ice crystals, although the current detection method also enables specific measurements above water saturation. For analysis the temperature and relative humidity in the chamber are set to the desired conditions, and the evolution of ice crystals on the wafer is monitored for 100 s, which is sufficient to activate all INPs. Due to the viscous flow conditions present in the FRIDGE chamber, the adjusted RH field may be disturbed after the first ice nucleation event and subsequent ice nucleation may occur at different RH. Since quantification of this effect is challenging and its magnitude/importance for the given setup is not entirely clear, we report all ice nucleation events at the same RH.

The primary cause of image misinterpretation that existed in earlier FRIDGE measurements stemmed from water droplets being miscounted as ice crystals. This was somewhat a result of overly sensitive ST and BT parameters; but more significantly, studies targeting hygroscopic marine aerosol samples have made plain that such particles grow by condensation and deliquescence at lower-than-expected vapor saturations. Samples heavily loaded with marine aerosol showed unreasonably high levels of detected INPs using the previous analysis technique, suggesting that hygroscopic aerosols were counted instead of INPs. Furthermore, a reanalysis of archived data and image files demonstrates that many counted objects did not grow or change brightness over time as would be expected for ice crystals in an icesupersaturated environment. Rather, many particles tended to activate immediately upon access to water vapor and maintain a constant size throughout the experiments (Fig. 2). An amplification factor that was applied within the software to make small brightness changes visible to the observer further masked the error by limiting the range of distinguishable grey levels. In contrast to small hygroscopic particles, ice crystals appear as defined objects that grow rapidly and steadily in the ice-supersaturated regime and ultimately result in signals



Figure 2. Mean object size in pixels recorded in 10 s increments, for a strongly ice nucleating silver iodide sample (blue) and a separate hygroscopic NaCl salt sample (red) vs. the measurement time. Measurements made at -18 °C and RH_{*i*} = 118 %.

with size and brightness magnitudes larger than hygroscopic particles.

The effect is most easily recognized by examining particle growth originating from idealized aerosol particle samples. Figure 2 shows the mean diameter in pixels of counted objects using the outdated image processing algorithm for a hygroscopic salt sample (red line) and an efficient ice nucleating silver iodide sample (blue line). Shortly after the introduction of the water vapor the first signal appears in the salt sample, but as time elapses the mean size of the counted objects remains constant at seven pixels. Contrastingly, the silver iodide sample shows a steady increase to a mean size of 600 pixels after 50 s. The phase of the objects counted by the algorithm can also be unraveled by analyzing the yield of detected objects in samples of hygroscopic and non-hygroscopic aerosols as a function of supersaturation with respect to ice and water (Fig. 3). This is done in Fig. 3, where the number of counted objects is plotted vs. relative humidity for both sea salt (a, b – hygroscopic) and silver iodide (c, d – non-hygroscopic, ice nucleating) at three temperature conditions. The count number from the sea salt sample does not depend on ice supersaturation and even decreases slightly with colder temperatures. However, when the data are re-plotted vs. water saturation, a correlation is observed, suggesting that there is a hygroscopic growth effect for liquid water. In contrast a strong exponential increase in count number as a function of ice supersaturation is visible in the silver iodide data.

Raw images of the wafers also show the clear difference between growing ice particles and small, stationary particles with microscopic amounts of surface-bound water (Figs. 4 and 5).

Repeated laboratory studies with well-known ice-active aerosolized materials including silver iodide, Snomax[®] and Arizona Test Dust consistently show the strong response of



Figure 3. Number of counted INPs on wafers with deposited sea salt (a, b) and silver iodide (c, d) particles vs. relative humidity with respect to ice (left panels) and water (right panels) for $-8 \degree C$ (red), $-13 \degree C$ (blue) and $-18 \degree C$ (black).



Figure 4. Unprocessed images of sample wafers 50 s after water vapor injection at T = -18 °C, RH_i = 118 % for (**a**) silver iodide and (**b**) NaCl.

these particles as INPs and their rapid growth, as illustrated in Figs. 2, 3, and 4.

To verify that small, slow-growing ice crystals are not neglected in the counting scheme, several experiments were performed with higher-resolution optical methods. Both a light microscope and a high-resolution lens (Navitar 12X



Figure 5. Microscope image of a sample laden with many small salt particles (upper left) and a single silver iodide particle (lower right). The hygroscopic salt facilitates the formation of many small spherical objects, while a distinct ice crystal grows on the sector with the deposited silver iodide. T = -18 °C, RH_i = 118 %.

Zoom) were utilized to observe in detail small subsections of wafer substrates in real time as ice crystals and water droplets formed and evolved (Fig. 5). Figure 5 shows a microscope image of a wafer with many small hygroscopic salt paticles and a single ice-active silver iodide particle sampled in separate sectors. The line demarcating the separation between microscopic droplets and a large ice crystal is clear.

Light polarization was also used to distinguish liquid isotropic droplets from ice crystals. The anisotropy of ice crystals ensures that reflected and refracted light intensity and polarization are a function of the incident parameters and geometry of the ice crystal (Thomson et al., 2009). Thus a polarization filter can be used to distinguish ice particles from unfrozen droplets (Fig. 6).

More high-resolution imaging was used to examine the sub-pixel-scale behavior of ice and droplet activation. Figure 7 shows the antiquated FRIDGE pixel domains (red grid) overlaying a high-resolution image. The shaded stars illustrate the grey level change from the reference picture if such a difference exists, with the white stars indicating the changes which exceeded the \mathcal{BT} . The plot demonstrates two important aspects of the image processing. First, \mathcal{BT} and \mathcal{ST} must be chosen to allow objects to overlap pixel boundaries. In Fig. 7 the bright spherical object with a diameter of 18 µm (growing from $6 \mu m$ before the measurement) is smaller than even a single pixel, yet it causes seven neighboring pixels to exceed \mathcal{BT} and therefore could be falsely counted as an ice crystal if $ST \leq 7$ pixels. Secondly, it highlights the critical nature of the density of sampling. Clearly, too much particle loading on the sample will cause overlapping signals, but too few particles will make statistical analyses more difficult (cf. Sect. 5). As an operational protocol we recommend investigators use initial wafer sampling followed by quick analysis



Figure 6. Successive microscope images of a sample substrate (a) without using a polarization filter and (b) with the filter inserted. The latter highlights the use of polarized light to identify frozen ice (I) vs. a liquid droplet (II), whereby the birefringence of the ice crystal sufficiently alters the light polarization to allow it to pass through the filter.

to determine estimated INP concentrations. Such initial results can be used to approximate appropriate sampling volumes (times).

These findings have led to alterations of the image processing procedure. First, \mathcal{BT} and \mathcal{ST} are carefully chosen to maximize the signal-to-noise ratio. Second, objects which are observed not to grow in ice saturated conditions are not counted as ice. Establishing these restrictions for FRIDGE measurements means that as much as 99% of previously counted objects are now neglected. As a consequence sampling volume, analysis temperatures and relative-humidity conditions must be altered in order to obtain statistically meaningful INP counts on a substrate.



Figure 7. Overlay of high-resolution substrate image taken using a Navitar $12 \times \text{zoom}$ lens and the outdated FRIDGE image processing (red grid cells = 1 Px, stars signal with which color those pixels would have been displayed).

4 Particle losses

Previously, the FRIDGE electrostatic sampling unit collection efficiency has been reported as near 100% over a wide range of sizes measured using a TSI 3936 Scanning Mobility Particle Sizer (SMPS; cf. Fig. 5; Klein et al., 2010a). More recently it is has been suggested that 100% retention efficiency, which was measured by Klein et al. (2010a), does not necessarily correspond to 100% deposition onto the sampling wafers. With the PEAC7 a number of tests have been run to ascertain potential particle losses within the collector housing and to determine the fraction of particles deposited onto the wafers.

In order to assess the deposition efficiency of the electrostatic collector, tracer fluorescein natrium particles (Carl Roth Fluorescein-Natrium (C.I. 45350), article number 5283.1) were aerosolized by dry dispersion in compressed air and simultaneously measured in parallel using filter sampling (47 mm Fluoropore filters, 0.2 µm pore size, Merck Millipore Ltd.) and the PEAC7. In parallel with sampling, the size spectra of the generated aerosols were determined using a TSI 3330 Optical Particle Sizer (OPS). After sampling, the material is washed from the respective substrates by submerging in 25 mL deionized ultrasonic water baths. The washing water is analyzed for the concentration of dissolved fluorescein sodium using a HACH DR/2010 spectrophotometer. The fluorescein mass concentration can be related to the 494 nm absorption maximum using a calibration for concentration as a function of absorbance. From the measured absorbances a deposition efficiency \mathcal{E} can be calculated that corresponds to the ratio of material absorbed onto the wafer M_{waf} vs. the material absorbed by the filter M_{filt} :

$$\mathcal{E} = \frac{M_{\rm waf}}{M_{\rm filt}},\tag{1}$$



Figure 8. Measured deposition efficiency in a PEAC7 aerosol sampler for 35 samples, exhibiting a range of fluorescein concentrations (lower abscissa scale, blue diamonds) and mean particle diameters (upper abscissa scale, red dots).

where the collection efficiency of the filter was verified to be unity using downstream sampling. Washing water was also used to irrigate the sample tubing to test for losses elsewhere in the collection system. Those tests were negative for fluorescein, and thus it is concluded that all particle losses occur within the FRIDGE wafer housing unit.

Figure 8 depicts measured deposition efficiencies for a range of fluorescein concentrations and varied particle size distributions expressed using the measured mean particle diameter. Mean particle diameters represent the averages of broad distributions spanning more than an order of magnitude of particle size as resolved by an OPS (0.3-10 µm scanning range). These systematic measurements show no correlation between either particle size or fluorescein concentration with deposition efficiency. By extension the combined measurements allow us to conclude for this experiment that deposition efficiency will be similarly independent of number concentration and particle size in the $0.5-3 \,\mu m$ range. In fact one advantage of FRIDGE is that it can be used to sample the entire aerosol size spectrum without artificial size cutoffs introduced by the sampling and/or measurement method. From the aggregate of experiments a mean collection efficiency $\overline{\mathcal{E}} \pm \Delta \mathcal{E} = 0.596 \pm 0.018$ is calculated, where $\Delta \mathcal{E}$ represents the standard error of the mean. Thus the particles deposited onto the FRIDGE wafers represent approximately 60% of the total ambient particles.

A second test of the collection efficiency with aerosolized mineral hematite particles was undertaken by scanning electron microscopy (SEM). The results from direct particle counting in the SEM were used as a verification test of the fluorescein results. Hematite was chosen as reference particles because its iron signature is easily identifiable using SEM and because the aerosolized hematite sample includes smaller particles (size distributions centered around $0.8-1.0 \,\mu\text{m}$) that are not as abundant in the fluorescein sam-

Wafer	OPS total particles	SEM counted particles		s SEM tota	SEM total particles		Efficiency [%]	
	Bounds:	Lower	Upper	Lower	Upper	Lower	Upper	
1*	177 132	162	386	63 302	150 832	35.7	85.2	
2	214 292	332	403	129731	157 475	60.5	73.5	
3	211 232	279	298	109 021	116 445	51.6	55.1	
				Mean efficiencies [%]:		49.3	71.3	
						60.3		

Table 1. Particle data for SEM measurements.

* The heavy clustering observed in wafer 1 was subsequently minimized by using a cyclone separator to remove the large particle fraction.

ples. Using the SEM's high-resolution imaging, 33 crosssectional scans of three wafers with precipitated hematite were made. Each scan consisted of 33 images, each covering 3.7×10^{-3} mm², thus resulting in total scanned areas of approximately 4 mm² per wafer. The scanned particle numbers were scaled by the areal ratio, assuming that the random SEM sampling procedure was representative of the total particle distribution. Thus the total particle number calculated from the SEM was compared with the particle number measured from the same aerosol using the TSI 3330 OPS to compute a mean efficiency of 60.3 %. The data presented in Table 1 illustrate that some uncertainty exists due to clustering of particles - lower and upper bounds correspond respectively to the assumptions that deposited agglomerates did or did not exist in the aerosol and thus were or were not present in the OPS counting. It is unknown if and how many clusters are present in the raw aerosol; however, some amount of clustering can be expected simply due to the deposition mechanism. Given the uncertainty and repetitive nature of individual particle counting, the excellent agreement with the fluorescein is remarkable.

Thus for the PEAC7 collector the identified mean collection efficiency $\overline{\mathcal{E}} \pm \Delta \mathcal{E} = 0.596 \pm 0.018$ and associated uncertainty must be considered when reporting measured INP concentrations as outlined in Sect. 5. Although deposition efficiencies are not expected to vary significantly for other collectors, or collectors run under unusual collection conditions, investigators should be aware of potential variability.

5 Experimental repeatability and analysis uncertainty

Due to the time-consuming nature of routine FRIDGE analysis it is impossible to repeat individual wafer analysis to the extent that an ensemble of nucleation counts would be generated for every wafer. Thus there exists some intrinsic uncertainty to FRIDGE wafer analysis, in that repeated cooling cycles on a single wafer will yield some spread in the number of observed INPs. Repetitive experiments have been conducted with exemplary wafers loaded with ambient atmospheric samples from the from the Taunus Observatory at Mt.

Kleiner Feldberg (826 m a.s.l., 50.221879° N, 8.446297° E), where daily INP sampling is conducted (see also Sect. 6). Through repeating such measurements for many samples, the weighted mean relative uncertainty in the counted absolute INP number is determined to be 20%. This reported uncertainty is taken from 226 individual measurements of 18 wafers measured with between 2 and 10 repetitions at 8 temperature and saturation conditions (T, $RH_i = -15 \circ C$, 110%; -20°C, 120%; -25°C, 119%; -25°C, 126%; -30 °C, 130 %; -30 °C, 132 %; -32 °C, 127 %; -32 °C, 134%) that cover the span of the conditions used in typical FRIDGE analyses. From the multiple analysis of each wafer at each saturation condition a mean INP number and standard deviation was determined. These standard deviations were used to calculate the relative error associated with each subset of repeated experiments (one wafer repeated n times represents one subset). Because no clear trend in relative uncertainty is observed with total wafer count number or saturation condition, we have sought to calculate a generally representative reproducibility envelope. Thus the relative errors of all 87 subsets were weighted by their mean INP and summed to arrive at the total weighted error. Three subsets which include a total of 30 measurements are shown in Fig. 9 to illustrate the spread of INP results for repeated samples. The details of the calculations and the entirety of the raw data are presented in the Supplement for the benefit of the community.

The reported relative error can be taken to be representative of the absolute count uncertainty for single measurements, and it is valid over the range of parameter space accessible to FRIDGE analysis. However, given the significant variability we encourage investigators to utilize the attached data set if they are interested in reprocessing the data for a narrower window of saturation conditions. At very low absolute INP numbers, 20 % can remain as a guidepost, but there are a host of cautionary issues, including the strong effect of single INPs. Furthermore, there also exists an upper limit for the resolvable INP number that is defined by the ability to distinguish individual ice domains vis-á-vis the image analysis protocols (Sect. 3). Thus an upper bound of sampling density must be defined at the user level. For typical conditions



Figure 9. Example data from wafer measurements repeated 10 times and reordered to resemble normal distributions. The weighted mean relative error determined from repeated measurements of 18 wafers (3–10 repetitions per wafer) is 20 %.

in the Frankfurt system average INP densities exceeding one INP per square millimeter become problematic.

In addition to the uncertainty based upon the replicability of INP counts on a given wafer, some added uncertainty stems from the FRIDGE collection units, which include flow regulators and, as outlined for the PEAC7, programmable, timed sampling. Thus for each unit and sampling protocol the uncertainty of the flow (ΔF) and the timing (Δt) must be empirically determined. These uncertainties combined with the INP count uncertainty and the uncertainty in collection efficiency (Sect. 4) can be used to calculate the individual measurement uncertainty using standard error propagation techniques. For example the ambient INP concentration IN/L can be calculated from the number of INPs counted on a single wafer:

$$IN/L = \frac{IN^{\#} - IN^{\#}_{blank}}{tF} \mathcal{E}^{-1},$$
(2)

where the INPs counted on a blank (clean) wafer under identical conditions $IN_{blank}^{\#}$ are subtracted from the INPs counted on the sample wafer $IN^{\#}$ and divided by the total volume sampled as determined by the product of the sampling time *t* and the sampling flow rate *F*, modulo the collection efficiency \mathcal{E} . Thus the uncertainty in the concentration can be expressed as

$$\Delta IN/L = |IN/L| \sqrt{\frac{\left(\Delta IN^{\#}\right)^{2} + \left(\Delta IN_{\text{blank}}^{\#}\right)^{2}}{\left(IN^{\#} - IN_{\text{blank}}^{\#}\right)^{2}} + \left(\frac{\Delta t}{t}\right)^{2} + \left(\frac{\Delta F}{F}\right)^{2} + \left(\frac{\Delta \mathcal{E}}{\mathcal{E}}\right)^{2}}, \quad (3)$$

$$\Delta IN^{\#} = 0.2 (IN^{\#}) \text{ and } \Delta IN^{\#}_{blank} = 0.2 (IN^{\#}_{blank}).$$
 (4)

In general the number of INPs counted on blank wafers is small or zero, and thus there is little or no contribution to the uncertainty from clean wafers. However, when analyzing wafers at low temperatures $T \leq -30$ °C, experimenters should cautiously verify the background detection from cleaned wafers. Similar error propagation procedures can be followed when calculating other quantities from INP counts measured using the FRIDGE system.

Typical laboratory and field measurements of INPs use repeated sampling schemes in order to collect broad data sets. Although laboratory methods may yield constant INP production, it is unlikely that natural systems will have less variability than the uncertainty intrinsic to the FRIDGE method that is reported here. Rather, observations to this point show that the methods uncertainty is significantly less than the natural variability captured by repeated measurements. Thus, it is a general recommendation that when reporting data both the intrinsic uncertainty and any statistical uncertainties associated with averaging, or the like, be independently reported.

5.1 The effects of wafer transport and storage

One significant advantage of the FRIDGE system is the ease of sample collection, wafer transport and storage, which makes it suited as a platform for a network of sampling stations. In general and for the results presented herein FRIDGE wafers are stored and transported in sealed wafer holders at ambient conditions. Systematic measurements of multiple wafers sampled from a single laboratory aerosol show little variation with storage time as depicted in Fig. 10. The samples were collected from test aerosols provided at the Aerosol Interaction and Dynamics in the Atmosphere (AIDA) test facility in Karlsruhe, Germany, during the Fifth International Ice Nucleation Workshop (FIN02; Möhler et al., 2015). In Fig. 10 the activated fractions, or the ratio of INP to total particle number determined using condensation particle counters, for wafers with short storage times are plotted vs. activated fractions from identical aerosol sampling but longer wafer storage times. Individual points represent the mean values of multiple wafers, with the error bars corresponding to the calculated standard deviations. For comparison the 1:1 line is plotted straddled by the $\pm 20\%$ envelope that is indicative of the reproducibility given repeated analysis of single wafers (cf. Sect. 5). The plot is illustrative in myriad ways. First it makes clear that variability encapsulated by wafer-to-wafer differences significantly outstrips the uncertainty in the measurement technique. For example, compare the error bars with the 20% envelope that is similar in span to the symbol size. Second, to within the uncertainty of the measurement there appears to be little or no effect of storage. It should be noted that for these wafers no special storage precautions were taken (stored in PetriSlides at ambient lab conditions). There may be materials (e.g., primary biological particles) for which storage time and/or conditions do play a role, and as such investigators should remain cautious. As expected the agreement is worse for the smallest activated



Figure 10. Effect of storage on sample integrity. Scatterplot of the fraction of INPs activated in aerosol samples that were analyzed shortly after sampling (abscissa) vs. the same parameter for samples from the same aerosol that were analyzed after storage (ordinate). Symbols correspond to different test aerosols as indicated, while the color codes represent the average storage time difference between analyses. The test aerosols were provided during the FIN02 workshop by the organizers (Möhler et al., 2015).

fractions, or fewest INPs, where the consequences of small changes in particle counts are amplified.

Repeated measurements at varying storage times make manifest the suitability of FRIDGE for monitoring applications and serve to reinforce the importance of clear and systematic approaches to the discussion of methodological uncertainties.

6 Re-evaluating FRIDGE using Saharan dust as a test case

On 16 April 2015 a Saharan dust event was sampled using the FRIDGE electrostatic deposition wafers at the Taunus Observatory at Mt. Kleiner Feldberg. For completeness, a description of the dust event is included within the Supplement. Nine samples were taken over $\approx 6 \,\text{h}$ that day and were subsequently analyzed at -16, -18, -20, -22 and -24 °C at ice supersaturations straddling water saturation. In Fig. 11 the INP concentrations measured at $RH_{water} = 101$ % are presented and compared with INP values calculated from the empirically based immersion freezing parameterization model developed by DeMott et al. (2010) that has since been adapted specifically to mineral dust (DeMott et al., 2015). The latter were calculated using the aerosol size spectra concurrently measured at the site with a TSI 3330 OPS and the nucleation temperature measured in kelvin T_k as input parameters. The DeMott et al. (2015) mineral dust parameteri-



Figure 11. Measured INP concentrations vs. calculated INP concentrations for mineral-dust-loaded aerosol transported to Mt. Kleiner Feldberg in April 2015. Error bars indicate the complete measurement uncertainty as enumerated in Eq. (3). The three lines correspond to three realizations of the DeMott et al. (2015) parameterization given three different calibration factors (Eq. 5). Where they do not appear, error bars are subsumed by the points.

zation,

$$n_{\rm INP}(T_{\rm k}) = ({\rm cf})(n_{a>0.5\,\mu\rm{m}})^{(\alpha(273.16-T_{\rm k})+\beta)} \exp(\gamma(273.16-T_{\rm k})+\delta),$$
(5)

is based on continuous-flow diffusion chamber (CFDC) measurements of both laboratory and naturally generated INPs and predicts INP concentration $n_{\text{INP}}(T_k)$ as a function of the total particle number concentration with diameters greater than 0.5 μ m $n_{a>0.5}$ μ m (cm⁻³) and the aforementioned temperature. The parameters cf, $\alpha = 0$, $\beta = 1.25$, $\gamma = 0.46$ and $\delta = -11.6$ are empirically determined, with cf segregated to account for instrument-specific calibration factors. For laboratory data cf = 1, while for atmospheric data cf = 3 is shown to yield a better parametric fit. As shown in Fig. 11, the FRIDGE data are best fit by the model using cf =0.636, while maintaining $\alpha = 0, \beta = 1.25, \gamma = 0.46$ and $\delta =$ -11.6. It is impossible to determine whether the observed deviation from the earlier parametric fits results from instrumental differences or from the fact that the DeMott et al. (2015) parameterization was developed using measurements subject to size cutoffs ranging from 1.5 to 2.4 µm and saturation conditions $RH_{water} = 105 \%$ (DeMott et al., 2010, 2015) that differed slightly from the FRIDGE measurements. It is reasonable to assume that some combination of these effects yields the measured trend. However, given these differences the overall agreement with the DeMott et al. (2015) model is very strong over a broad range of temperature and more than 3 orders of magnitude of INP concentration.

The agreement of FRIDGE measurements with the De-Mott et al. (2015) parameterization that is based on a host of data collected from both laboratory and field measurements for a variety of mineral dusts is not in itself proof that FRIDGE measurements yield strictly constrained absolute INP numbers. However, it is important evidence that the uncertainties of the FRIDGE instrument are well enough constrained such that the measurements can be systematically compared with other reported INP data. Furthermore, De-Mott et al. (2015) demonstrated their simple model to be in agreement with the more complex surface-area-based parameterization of Niemand et al. (2012). Niemand et al. (2012) utilized data from cloud chamber expansion experiments performed at the AIDA facility to construct an immersion freezing parameterization, whereby temperature and dust particle surface area are used to calculate INP concentration. Thus the FRIDGE results are transitively linked with the experimental results from AIDA and therefore compare favorably with multiple natural and experimental systems.

7 Conclusions

The FRIDGE electrostatic deposition and wafer analysis system for INPs is a utilitarian tool for INP collection and evaluation that possesses several unique advantages. Collection units themselves are compact and simple to operate, and analysis for INP counting and particle characterization vis-á-vis scanning electron microscopy can be done in controlled laboratory settings. For these reasons a network of FRIDGE sampling and analysis instruments has begun to emerge, targeting globally distributed long-term measurements. The PEAC7 we have described here is a remotely controlled collector that can be automated for intermittent sampling and has been designed to target monitoring applications. FRIDGE can also complement other INP measurement techniques like continuous-flow diffusion chambers and droplet freezing arrays. For example, FRIDGE can be used to measure and identify larger supermicron particles that may be a significant component of INPs in some environments (Mason et al., 2016).

Challenges for the FRIDGE system include technical limitations with regards to the method's temporal resolution and a previously flawed analysis system. The former challenge can and will be incrementally addressed to improve the sampling technique and chain of analysis. While those incremental changes will potentially improve spatial and temporal resolution, they will not affect the underlying measurement principle and/or uncertainties.

Herein we have addressed the latter challenge by laying bare where problems have occurred in the past and how those problems can be systematically addressed to obtain accurate INP number results. First and foremost it is clear that a majority of counting error can result from the choice of image analysis parameters. Although the suggested values for image thresholds are not failsafe, they should be taken as a guide, and any FRIDGE user should be aware of the potential pitfalls detailed in Sect. 3. The FRIDGE collection system has also been re-examined, and it has been determined that deposition to non-wafer parts of the collector housing account for particle losses within the system. Experiments utilizing fluorescein tagged aerosol particles determined a mean deposition collection efficiency of $\overline{\mathcal{E}} \pm \Delta \mathcal{E} =$ 0.596 ± 0.018 , a value which is supported by direct SEM counting of hematite particles. Similarly, repeated analysis of single wafers has helped to constrain the uncertainty of any single INP number measurement to be ± 20 %. These factors plus any additional uncertainty in measurement duration and volume must be considered in order to construct a complete treatment of FRIDGE measurement uncertainty. Under most experimental conditions the repeatability will dominate other uncertainties, and thus $\pm 20\%$ is a useful guidepost for the intrinsic measurement uncertainty. However, as illustrated in Figs. 10 and 11, the natural variability typically observed outweighs the intrinsic uncertainty. Thus it is suggested that any communicated FRIDGE data should independently specify both this intrinsic measurement uncertainty and any statistical uncertainty associated with measurements whereby multiple sampling is used to illuminate variability.

A test case of the re-evaluated FRIDGE system illustrates agreement with other INP measurement methods that utilize both laboratory and natural aerosols. We find agreement up to -16 °C with the parameterization previously developed by DeMott et al. (2015) that uses aerosol number concentration to predict INP concentration. Thus the FRIDGE observations further support the DeMott et al. (2015) hypothesis that laboratory results for ice nucleation in idealized systems have applicability to natural atmospheric aerosols. Although such parameterizations have begun to be used in prognostic ice nucleation schemes for cloud and climate models (Fan et al., 2014a, b), the FRIDGE work also makes clear that among other things, investigations must continue to strive for a deeper understanding of instrument-dependent parametric variation.

Significant resources and energy are currently focused on addressing the broad issues that surround intercomparison of data collected using different techniques. In this regard FRIDGE has taken part in a series of instrument intercomparisons, beginning in March 2015 with the FIN02 campaign at the AIDA facility (Möhler et al., 2015). There more than a dozen different ice-nuclei-counting instruments and techniques were collected to measure and compare various iceactive materials in a rigorously designed laboratory environment. As a follow-up FRIDGE joined the third component of the intercomparison (FIN03) in September 2015, where measurements focused on utilizing multiple INP sampling techniques at a single field site (Storm Peak Laboratory, Steamboat Springs, CO, USA). Data analysis and synthesis from these campaigns is ongoing and will be available in upcoming publications. These efforts are expected to be a big step forward for the scientific understanding of how best to measure, characterize and systematically compare ice nuclei.

The Supplement related to this article is available online at doi:10.5194/amt-9-1313-2016-supplement.

Acknowledgements. Support and funding for this work were contributed by the Deutsche Forschungsgemeinschaft (DFG) under the Research Unit FOR 1525 (INUIT), the EU FP7-ENV-2013 BACCHUS project under Grant Agreement 603445, the Swedish Research Council, the Swedish Research Council FORMAS and the Nordic Top-Level Research Initiative CRAICC. The provision of test aerosols by the FIN02 organizing team is gratefully acknowledged.

Edited by: B. Ervens

References

- Ansmann, A., Atkinson, J., Decesari, S., Bühl, J., Facchini, M. C., Gysel, M., Baltensperger, U., Herrmann, E., Connolly, P., Crooks, M., Rosenfeld, D., and Sierau, B.: Definition of the BACCHUS aerosol/cloud database (structure, contents), first preliminary data set (some cloud products from satellite observations and first CCN/IN data), Deliverable report, TROPOS, Leipzig, Germany, 18 pp., 2014.
- Ardon-Dryer, K. and Levin, Z.: Ground-based measurements of immersion freezing in the eastern Mediterranean, Atmos. Chem. Phys., 14, 5217–5231, doi:10.5194/acp-14-5217-2014, 2014.
- Ardon-Dryer, K., Levin, Z., and Lawson, R. P.: Characteristics of immersion freezing nuclei at the South Pole station in Antarctica, Atmos. Chem. Phys., 11, 4015–4024, doi:10.5194/acp-11-4015-2011, 2011.
- Bingemer, H., Klein, H., Ebert, M., Haunold, W., Bundke, U., Herrmann, T., Kandler, K., Müller-Ebert, D., Weinbruch, S., Judt, A., Wéber, A., Nillius, B., Ardon-Dryer, K., Levin, Z., and Curtius, J.: Atmospheric ice nuclei in the Eyjafjallajökull volcanic ash plume, Atmos. Chem. Phys., 12, 857–867, doi:10.5194/acp-12-857-2012, 2012.
- Born, J., Möhler, O., Haunold, W., Schrod, J., Brooks, I., Norris, S., Brooks, B., Hill, M., and Leisner, T.: Meteorological and Aerosol Sensing with small Unmanned Aerial Systems, in: EGU General Assembly Conference Abstracts, edited by: Abbasi, A. and Giesen, N., vol. 14 of EGU General Assembly Conference Abstracts, Vienna, AT, 22–27 April 2012, p. 897, 2012.
- Bundke, U., Nillius, B., Jaenicke, R., Wetter, T., Klein, H., and Bingemer, H.: The fast Ice Nucleus chamber FINCH, Atmos. Res., 90, 180–186, doi:10.1016/j.atmosres.2008.02.008, 2008.
- DeMott, P. J., Prenni, A. J., Liu, X., Kreidenweis, S. M., Petters, M. D., Twohy, C. H., Richardson, M. S., Eidhammer, T., and Rogers, D. C.: Predicting global atmospheric ice nuclei distributions and their impacts on climate, P. Natl. Acad. Sci. USA, 107, 11217–11222, doi:10.1073/pnas.0910818107, 2010.
- DeMott, P. J., Möhler, O., Stetzer, O., Vali, G., Levin, Z., Petters, M. D., Murakami, M., Leisner, T., Bundke, U., Klein, H., Kanji, Z. A., Cotton, R., Jones, H., Benz, S., Brinkmann, M., Rzesanke, D., Saathoff, H., Nicolet, M., Saito, A., Nillius, B.,

Bingemer, H., Abbatt, J., Ardon, K., Ganor, E., Georgakopoulos, D. G., and Saunders, C.: Resurgence in ice nuclei measurement research, B. Am. Meteorol. Soc., 92, 1623–1635, doi:10.1175/2011BAMS3119.1, 2011.

- DeMott, P. J., Prenni, A. J., McMeeking, G. R., Sullivan, R. C., Petters, M. D., Tobo, Y., Niemand, M., Möhler, O., Snider, J. R., Wang, Z., and Kreidenweis, S. M.: Integrating laboratory and field data to quantify the immersion freezing ice nucleation activity of mineral dust particles, Atmos. Chem. Phys., 15, 393–409, doi:10.5194/acp-15-393-2015, 2015.
- Fan, J., Leung, L. R., DeMott, P. J., Comstock, J. M., Singh, B., Rosenfeld, D., Tomlinson, J. M., White, A., Prather, K. A., Minnis, P., Ayers, J. K., and Min, Q.: Aerosol impacts on California winter clouds and precipitation during CalWater 2011: local pollution versus long-range transported dust, Atmos. Chem. Phys., 14, 81–101, doi:10.5194/acp-14-81-2014, 2014a.
- Fan, J., Leung, L. R., DeMott, P. J., Comstock, J. M., Singh, B., Rosenfeld, D., Tomlinson, J. M., White, A., Prather, K. A., Minnis, P., Ayers, J. K., and Min, Q.: Corrigendum to Aerosol impacts on California winter clouds and precipitation during Cal-Water 2011: local pollution versus long-range transported dust published in Atmos. Chem. Phys., 14, 81–101, 2014, Atmos. Chem. Phys., 14, 3063–3064, doi:10.5194/acp-14-3063-2014, 2014b.
- Hallett, J. and Mossop, S. C.: Production of secondary ice particles during the riming process, Nature, 249, 26–28, doi:10.1038/249026a0, 1974.
- Jiang, H., Yin, Y., Su, H., Shan, Y., and Gao, R.: The characteristics of atmospheric ice nuclei measured at the top of Huangshan (the Yellow Mountains) in Southeast China using a newly built static vacuum water vapor diffusion chamber, Atmos. Res., 153, 200– 208, doi:10.1016/j.atmosres.2014.08.015, 2015.
- Klein, H., Haunold, W., Bundke, U., Nillius, B., Wetter, T., Schallenberg, S., and Bingemer, H.: A new method for sampling of atmospheric ice nuclei with subsequent analysis in a static diffusion chamber, Atmos. Res., 96, 218–224, doi:10.1016/j.atmosres.2009.08.002, 2010a.
- Klein, H., Nickovic, S., Haunold, W., Bundke, U., Nillius, B., Ebert, M., Weinbruch, S., Schuetz, L., Levin, Z., Barrie, L. A., and Bingemer, H.: Saharan dust and ice nuclei over Central Europe, Atmos. Chem. Phys., 10, 10211–10221, doi:10.5194/acp-10-10211-2010, 2010b.
- Lange, M., Argyrides, M., Ioannou, S., Keleshis, C., and Levin, Z.: Unmanned Aerial Systems as Versatile Tools for Atmospheric and Environmental Research, in: EGU General Assembly Conference Abstracts, vol. 15 of EGU General Assembly Conference Abstracts, Vienna, AT, 7–12 April 2013, p. 7851, 2013.
- Mason, R. H., Si, M., Chou, C., Irish, V. E., Dickie, R., Elizondo, P., Wong, R., Brintnell, M., Elsasser, M., Lassar, W. M., Pierce, K. M., Leaitch, W. R., MacDonald, A. M., Platt, A., Toom-Sauntry, D., Sarda-Estève, R., Schiller, C. L., Suski, K. J., Hill, T. C. J., Abbatt, J. P. D., Huffman, J. A., DeMott, P. J., and Bertram, A. K.: Size-resolved measurements of ice-nucleating particles at six locations in North America and one in Europe, Atmos. Chem. Phys., 16, 1637–1651, doi:10.5194/acp-16-1637-2016, 2016.
- Möhler, O., Cziczo, D. J., DeMott, P. J., Hiranuma, N., and Petters, M. D.: The 5th International Ice Nucleation Workshop Activities FIN-1 and FIN-2: Overview and selected results, AGU Fall Meeting, 14–18 December 2015, San Francisco, 2015.

- Niemand, M., Möhler, O., Vogel, B., Vogel, H., Hoose, C., Connolly, P., Klein, H., Bingemer, H., DeMott, P., Skrotzki, J., and Leisner, T.: A particle-surface-area-based parameterization of immersion freezing on desert dust particles, J. Atmos. Sci., 69, 3077–3092, doi:10.1175/JAS-D-11-0249.1, 2012.
- Schrod, J., Bingemer, H., Haunold, W., Curtius, J., Decesari, S., Marinoni, A., Rinaldi, M., Bonasoni, P., and Cristofanelli, P.: Ice nuclei measurements at a high altitude remote station in the Northern Apennines, in: EGU General Assembly Conference Abstracts, vol. 15 of EGU General Assembly Conference Abstracts, Vienna, AT, 7–12 April 2013, p. 1925, 2013.
- Thomson, E. S., Wilen, L. A., and Wettlaufer, J. S.: Light scattering from an isotropic layer between uniaxial crystals, J. Phys.-Condens. Mat., 21, 195407, doi:10.1088/0953-8984/21/19/195407, 2009.
- Vali, G., DeMott, P. J., Möhler, O., and Whale, T. F.: Technical Note: A proposal for ice nucleation terminology, Atmos. Chem. Phys., 15, 10263–10270, doi:10.5194/acp-15-10263-2015, 2015.